

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 30 2009

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130142</u>	
Name <u>Scott E. Gardner</u> ID# <u>6222</u> Date <u>24 Jun 09</u>	
A	Agency <u>Directorate of Emergency Services</u> Phone # <u>384-3888</u>
Instrument Location <u>Bldg #656, A St, Ft Richardson, AK 99505</u>	
B	Alco S/N <u>78820</u> Target Value <u>.079</u> High Pressure <u>700 psi</u>
Alco Test Values 1 <sup>st</sup> Alco <u>.083</u> 2 <sup>nd</sup> Alco <u>.082</u>	
Signature <u>Scott E. Gardner</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130142 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142 ✓

JUNE 24, 2009

OPERATOR'S NAME:

SARONER/SCOTT/E

OPERATOR'S NUMBER: 6222

SUBJECT'S LAST NAME:

N/A

SUBJECT'S FIRST NAME/MI :

N/A

O.L. #: N/A

DEPT/AGENCY: FRN4

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: 78820 ✓

### --- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.75 in  
ALCO TARGET .078 10:29  
BLANK TEST .000 10:30  
INTERNAL STANDARD VERIFIED 10:30  
ALCO TV 29.75 in .003 10:31  
BLANK TEST .000 10:31  
SUBJECT SAMPLE .000 10:32  
BLANK TEST .000 10:33  
ALCO TV 29.75 in .002 10:33  
BLANK TEST .000 10:34

*BMB 7/14/09*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

JUNE 24, 2009 ✓

TIME 10:35

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 44c  
BAROMETER: 29.75 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ0123456789  
abcdefghijklmnopqrstuvwxyz